Attachment D: Confined Space Meter Calibration Log

Directions: Confined Space meters must be "bump tested" before each use per the manufacturer's instructions. Confined Space meters must be periodically calibrated per the manufacturer's instructions. The person performing the bump test or calibration must be trained in the process. Each functional area will keep this log for at least one year to support Confined Space Program Evaluations.

Confined Space Meter Model &Serial #:_____

Date	Oxygen (O2)	Lower Explosive Limit (LEL)	Hydrogen Sulfide (H2S) or Toxic	Carbon Monoxide (CO)	Calibrated by: (Signature)